

Qualitative Compound Report

Analysis Info

Analysis Name D:\Data\ggc\LR1-113_P1-D-4_01_6912.d
Method lc_qtof_directospos.m
Sample Name LR1-113
Comment Sv: DCM + MeOH + HCOOH
Duran


Acquisition Date 3/6/2023 1:26:01 PM

Operator ELB
Instrument micrOTOF-Q II



Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	3.5 Bar
Focus	Not active	Set Capillary	4000 V	Set Dry Heater	200 °C
Scan Begin	100 m/z	Set End Plate Offset	-500 V	Set Dry Gas	7.0 l/min
Scan End	1500 m/z	Set Collision Cell RF	300.0 Vpp	Set Divert Valve	Source



Meas. m/z	#	Ion Formula	m/z	err [ppm]	mSigma	# mSigma	Score	rdb	e ⁻ Conf	N-Rule
320.1450	1	C18H20F2NO2	320.1457	2.0	7.0	1	100.00	8.5	even	ok