

Qualitative Compound Report

Analysis Info


Analysis Name D:\Data\ggc\LR2-187_P1-B-6_01_10423.d
 Method lc_ms directos tunelow291025.m
 Sample Name LR2-187
 Comment Sv: DCM + MeOH + HCOOH
 Fernando Durán

Acquisition Date 11/4/2025 12:13:06 PM

Operator EB - MV
 Instrument micrOTOF-Q

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	2.0 Bar
Focus	Not active	Set Capillary	4000 V	Set Dry Heater	200 °C
Scan Begin	100 m/z	Set End Plate Offset	-500 V	Set Dry Gas	7.0 l/min
Scan End	1200 m/z	Set Collision Cell RF	250.0 Vpp	Set Divert Valve	Source



Meas. m/z	#	Ion Formula	m/z	err [ppm]	mSigma	# mSigma	Score	rdB	e ⁻ Conf	N-Rule
410.1663	1	C26H21FN3O	410.1663	-0.1	4.6	1	100.00	17.5	even	ok