

Qualitative Compound Report

Analysis Info

Analysis Name D:\Data\ggc\24-051601_P1-E-5_01_9203.d
 Method lc_qtof_directospos.m
 Sample Name 24-051601
 Comment ZnNAM
 Sv: MeOH
 Evelina Gloria Ferrer

Acquisition Date 5/24/2024 12:10:27 PM

Operator MV
 Instrument micrOTOF-Q II

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	3.5 Bar
Focus	Not active	Set Capillary	4000 V	Set Dry Heater	200 °C
Scan Begin	70 m/z	Set End Plate Offset	-500 V	Set Dry Gas	7.0 l/min
Scan End	1000 m/z	Set Collision Cell RF	300.0 Vpp	Set Divert Valve	Source

