

# Qualitative Compound Report

## Analysis Info


Analysis Name D:\Data\ggc\CYN Ac\_P1-D-2\_01\_8389.d  
 Method lc\_qtof\_directospos.m  
 Sample Name CYN Ac  
 Comment Sv: DCM + MeOH + HCOOH  
 Garcia Linares

Acquisition Date 12/27/2023 10:41:05 AM

Operator ELB  
 Instrument micrOTOF-Q II

## Acquisition Parameter

|             |            |                       |           |                  |           |
|-------------|------------|-----------------------|-----------|------------------|-----------|
| Source Type | ESI        | Ion Polarity          | Positive  | Set Nebulizer    | 3.5 Bar   |
| Focus       | Not active | Set Capillary         | 4000 V    | Set Dry Heater   | 200 °C    |
| Scan Begin  | 70 m/z     | Set End Plate Offset  | -500 V    | Set Dry Gas      | 7.0 l/min |
| Scan End    | 1000 m/z   | Set Collision Cell RF | 300.0 Vpp | Set Divert Valve | Source    |



| Meas. m/z | # | Ion Formula | m/z      | err [ppm] | mSigma | # mSigma | Score  | rdb | e <sup>-</sup> Conf | N-Rule |
|-----------|---|-------------|----------|-----------|--------|----------|--------|-----|---------------------|--------|
| 411.1414  | 1 | C21H24NaO7  | 411.1414 | 0.1       | 1.1    | 1        | 100.00 | 9.5 | even                | ok     |