

Qualitative Compound Report

Analysis Info


Analysis Name D:\Data\ggc\BM79#1_P1-B-3_01_9658.d
 Method lc_ms directos tunelow300924.m
 Sample Name BM79#1
 Comment Sv: MeOH + HCOOH
 Sergio Szajnman

Acquisition Date 11/8/2024 12:54:38 PM

Operator EB - MV
 Instrument micrOTOF-Q

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	2.0 Bar
Focus	Not active	Set Capillary	4000 V	Set Dry Heater	200 °C
Scan Begin	100 m/z	Set End Plate Offset	-500 V	Set Dry Gas	7.0 l/min
Scan End	1000 m/z	Set Collision Cell RF	250.0 Vpp	Set Divert Valve	Source



Meas. m/z	#	Ion Formula	m/z	err [ppm]	mSigma	# mSigma	Score	rdb	e ⁻ Conf	N-Rule
553.2740	1	C33H42NaO4Si	553.2745	0.9	8.9	1	100.00	13.5	even	ok