

Qualitative Compound Report

Analysis Info


Analysis Name D:\Data\ggc\HG-3j2_P1-D-5_01_10045.d
 Method lc_ms directos tunelow300924.m
 Sample Name HG-3j2
 Comment Sv: MeOH
 Szajnman

Acquisition Date 2/14/2025 2:25:02 PM

Operator EB - MV
 Instrument micrOTOF-Q

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	2.0 Bar
Focus	Not active	Set Capillary	4000 V	Set Dry Heater	200 °C
Scan Begin	100 m/z	Set End Plate Offset	-500 V	Set Dry Gas	7.0 l/min
Scan End	1200 m/z	Set Collision Cell RF	250.0 Vpp	Set Divert Valve	Source



Meas. m/z	#	Ion Formula	m/z	err [ppm]	mSigma	# mSigma	Score	rdb	e ⁻ Conf	N-Rule
423.2317	1	C25H36NaO2S	423.2328	2.7	29.0	2	71.37	7.5	even	ok