

Qualitative Compound Report

Analysis Info


Analysis Name D:\Data\ggc\SJ334_P1-D-3_01_9160.d
 Method lc_qtof_directospos.m
 Sample Name SJ334
 Comment Sv: MeOH + HCOOH
 Szajnman

Acquisition Date 5/6/2024 4:03:25 PM

Operator ELB
 Instrument micrOTOF-Q II

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	3.5 Bar
Focus	Not active	Set Capillary	4000 V	Set Dry Heater	200 °C
Scan Begin	70 m/z	Set End Plate Offset	-500 V	Set Dry Gas	7.0 l/min
Scan End	1000 m/z	Set Collision Cell RF	300.0 Vpp	Set Divert Valve	Source



Meas. m/z	#	Ion Formula	m/z	err [ppm]	mSigma	# mSigma	Score	rdb	e ⁻ Conf	N-Rule
335.0863	1	C11H21NaO8P	335.0866	0.9	0.5	1	100.00	1.5	even	ok