

Qualitative Compound Report

Analysis Info


Analysis Name D:\Data\ggc\VL-80p_P1-B-7_01_9662.d
 Method lc_ms directos tunelow300924.m
 Sample Name VL-80p
 Comment Sv: MeOH + HCOOH
 Sergio Szajnman

Acquisition Date 11/8/2024 1:15:06 PM

Operator EB - MV
 Instrument micrOTOF-Q

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	2.0 Bar
Focus	Not active	Set Capillary	4000 V	Set Dry Heater	200 °C
Scan Begin	100 m/z	Set End Plate Offset	-500 V	Set Dry Gas	7.0 l/min
Scan End	1000 m/z	Set Collision Cell RF	250.0 Vpp	Set Divert Valve	Source



Meas. m/z	#	Ion Formula	m/z	err [ppm]	mSigma	# mSigma	Score	rdb	e ⁻ Conf	N-Rule
309.0413	1	C ₁₂ H ₁₄ NaO ₆ S	309.0403	-3.0	7.0	1	71.07	5.5	even	ok