

Qualitative Compound Report

Analysis Info


Analysis Name D:\Data\ggc\25-020506_P1-E-6_01_10054.d
 Method lc_ms directos tunelow300924.m
 Sample Name 25-020506
 Comment NBocDOPAsil
 Sv: DCM + MeOH + HCOOH
 Oscar Varela

Acquisition Date 2/21/2025 10:20:52 AM

Operator EB - MV
 Instrument micrOTOF-Q

Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	2.0 Bar
Focus	Not active	Set Capillary	4000 V	Set Dry Heater	200 °C
Scan Begin	100 m/z	Set End Plate Offset	-500 V	Set Dry Gas	7.0 l/min
Scan End	1200 m/z	Set Collision Cell RF	250.0 Vpp	Set Divert Valve	Source



Meas. m/z	#	Ion Formula	m/z	err [ppm]	mSigma	# mSigma	Score	rdb	e ⁻ Conf	N-Rule
504.2936	1	C25H47NNaO4Si2	504.2936	0.0	13.1	1	100.00	4.5	even	ok